

Certificate # LT/ROHS/2113

## RoHs Certificate of Compliance

This document certifies that following LABOMED Surgical Microscopes are RoHs compliant ,they do not contain the substances listed in the table below in concentrations exceeding the Maximum Control value (MCV)\*\* as per Directive 2015/863/EU in force by June 4th, 2015.

### Labomed Surgical Microscopes:

**Prima ENT, Prima DNT, Prima OPH,Prima Lite,Prima Pro & Prima swift**

Substances Maximum control Value (MCV)\*\*

Cadmium (0.01 %)

Lead (0.1 %)

Mercury (0.1 %)

Hexavalent chromium (0.1 %)

Polybrominated biphenyls (PBB) (0.1 %)

Polybrominated diphenyl ethers (PBDE) (0.1 %)

Bis(2-ethylhexyl) phthalate (DEHP) (0.1 %)

Butyl benzyl phthalate (BBP) (0.1 %)

Dibutyl phthalate (DBP) (0.1 %)

Diisobutyl phthalate (DIBP) (0.1 %)

**Reference:** Restriction of Hazardous Substances, RoHS.

- DIRECTIVE 2015/863/EU
- DIRECTIVE 2011/65/EU (2023-09-01)

### Exemptions:

**Annex Point # 6** Lead as an alloying element in steel containing upto .35% Lead by weight ,Aluminium containing up to 0.4% Lead by weight ,and as cooper alloy containing up to 4% Lead by weight.

**Annex Point # 7**Lead in ceramic parts

**Annex Point # 8** Cadmium and its compounds in electrical contacts

\*In order to validate compliance ,all products are evaluated to the homogenous material level.  
Authorized Signatory



Sukhvinder Bal  
(Manager QA & RA)  
E-mail: sbal@labomed.in

Date: Mar 1,2024